

# Terms of Reference

**Note:** Bidders must state in the Statement of Compliance either “**Comply**” or “**Not Comply**” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Unless otherwise categorically stated hereunder, evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent.

## A. General Scope of Work

No.	General Scope of Work	Statement of Compliance
1	<i>[Emergency Procurement Under Bayanihan To Recover As One Act (Bayanihan 2) - Renovation of Morgue Facility with Equipment Component (Design and Construct)]</i>	

## Detailed General Scope of Work

ITEM NO.	Scope of Work	STATEMENT OF COMPLIANCE
	<b>Renovation of Morgue Facility with Equipment Component (Design and Construct)</b>	
	<b>A. Scope of Work</b>	
	1. Mobilization/ Demobilization 2. Temporary Facilities 3. Safety and Health Protocols 4. Demolition Works (including hauling and disposal) 5. Masonry Works 6. Tile Works 7. Plumbing Works 8. Electrical Works (including lighting fixtures) 9. Ceiling Works 10. Painting Works 11. Concrete Counter Tile Finish 12. Cabinetry (under counter only) 13. Office Tables and Chairs 14. Supply and Installation of air-conditioning unit for the following areas: <ul style="list-style-type: none"> <li>• Autopsy Area (At least 1.5 HP Split-type inverter and exhaust)</li> <li>• Family waiting lounge (At least 1.0 HP Split-type inverter and exhaust)</li> <li>• Cadaver holding Area (At least 1.0 HP Split-type inverter and exhaust)</li> <li>• Cadaver Area (At least 1.0 HP Split-type inverter and exhaust)</li> </ul>	
	<b>B. Areas Requiring Renovation</b>	
	1. Family Waiting Lounge 2. Stock Room 3. Cadaver Freezer 4. Dirty Room 5. Cadaver Holding Room 6. Work Room 7. Wash Room 8. Clean Room 9. Toilet & Bath	
	<b>C. Mechanical Requirements/ Equipment</b>	
	<b>1. Mortuary Freezer/ Refrigerator Premium 6x3 (18 Cadavers)</b>	
	<b>Technical Specifications</b>	
	<ul style="list-style-type: none"> <li>• Dimensions (mm): 2500 x 1620 x 1800</li> <li>• Body Tray Size: 1940 x 580mm</li> <li>• Finished: Stainless Steel</li> <li>• Voltage: AC220v Single Phase</li> <li>• Total Power: 880W</li> <li>• Total Weight: 600 kg</li> <li>• Cryogenic Temperature: 0°C ~ - 35 °C</li> <li>• Insulating Layers: 10 centimeters</li> <li>• Refrigerant: R600A</li> <li>• Compressor: Separate Temperature Controller with six compressors (Individual Compressor)</li> <li>• Thermostat: Microcomputer Control Digital Display</li> <li>• Temperature Reaches: -15 to 5 Degrees Adjustable 18°C Within 24 Hours</li> <li>• Or approved equivalent</li> </ul>	
	<b>Design Features:</b>	
	<ul style="list-style-type: none"> <li>• Unit ships fully assembled, Plug and Play Installation.</li> <li>• Voltage 220v.</li> <li>• All quality embossed galvanized steel construction with a flush mounted door.</li> <li>• Heavy-duty chrome plated for long lasting use.</li> <li>• Provided with interior roller racks.</li> <li>• Removable racks for cleaning purposes.</li> <li>• Stainless Steel Body Trays, refrigerator wooden case for easy of transport.</li> <li>• Standard temperature range of storage corpses is from 0°C to -10°C, -15°C ~ 5°C adjustable.</li> </ul>	

	<ul style="list-style-type: none"> <li>• Each unit will have six (6) pieces high quality universal wheel to ensure mobility</li> <li>• Or approved equivalent</li> </ul>	
	<b>2. Autopsy Table</b>	
	<b>Technical Specifications:</b>	
	Product Size (mm): 2300x800x800mm Load Capacity: 200 kg Table Height: 80mm Water Bath Size: 380x320x150 Material: Stainless Steel With Lens With Tap With Hose With Spray Or approved equivalent	
	<b>3. Mortuary Lifter</b>	
	<b>Technical Specifications:</b>	
	Keyword: Stainless Steel Height: 1940x580mm System: Hydraulic with Universal Wheels Or approved equivalent	
	<b>4. Emergency Shower and Eye Wash Station</b>	
	<b>Technical Specifications:</b>	
	The emergency shower should deliver a pattern of water with a diameter of at least 50.8 cm (20 inches) at 152.4 cm (60 inches). The shower head be between 208.3 and 243.8 cm (82-96 inches) from the floor. The minimum volume of spray should be 75.7 liters/minute (20 gallons/minute) for a minimum time of 15 minutes.	
	The shower should also be designed so that it can be activated in less than 1 second, and it remains operational without the operator's hand on the valve (or lever, handle, etc.). This valve should not be more than 173.3 cm (69 inches) in height. If enclosures are used, ensure that there is an unobstructed area of 86.4 cm (34 inches) in diameter.	
	Water closet should be controlled by foot pedal.	
	<b>Service Warranty: 1 year on parts, 3 years on service</b>	
	<b>Free semestral preventive maintenance and calibration</b>	
	<b>INCLUSIONS:</b>	
	UPS (Uninterrupted Power Supply) for freezer	
	TRAINING FOR THE MACHINE	
	ONSITE TRAINING OF OUR BIOMEDICAL ENGINEERS AND TECHNICAL SPECIALIST	

Conforme:
Company/Bidder's Name:
Name and Signature of Authorized Representative:
Date: