

Item No.	Item	Specification	Quantity	Unit	Ceiling Price Per Unit	Total Amount
1	Nasal High Flow		1	unit	450,000.00	450,000.00
	<b><u>Clinical Features:</u></b>					
	Temperature Settings	31 - 37 C				
	Max Temperature	43 C				
	Humidity Requirement	> 12 - > 33 mg/L (10 - 95% Relative Humidity)				
	Flow Range	10 - 60 Lpm (minimum range)				
	Oxygen Output	60Lpm				
	FiO2 Monitoring Range	21 - 100%				
	FiO2 Monitoring Accuracy	+/- 4%				
	Warm - up Time	at least 10 - 30 mins.				
		( at temp range 31 - 37 C)				
	Alarms	< / = 45 dbA				
	Alarm Silence	At least 1 min.				
	<b><u>Physical</u></b>					
	Weight	less than 10 lbs.				
	Mobility	with stand and at least 4 wheels				
	<b><u>Electrical</u></b>					
	Frequency	50 - 60Hz				
	Current/Voltage Requirement	Autovolt (110 -240)				
	UPS (Uninterrupted Power Supply)	(at least 30 mins to 1 hour)				
	<b>Accessories</b>					
	Breathing Tubes	Disposable				
	Humidification Chamber	Disposable				
	HFNC Interface	Small, Medium, Large				
		<b>Documentary Requirements:</b>				
		1.Certificate of availability of stocks and can deliver within the prescribed delivery schedule.				
		2) Valid and current License to Operate (LTO) as medical device distributor issued by the Philippine Food and Drug Administration. In Case of expired LTO, the following must be submitted: i) Copy of expiredLTP, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO				
		<b>Requirements if awarded the Contract:</b>				

Item No.	Item	Specification	Quantity	Unit	Ceiling Price Per Unit	Total Amount
		1. <b>Delivery and Completion Schedule:</b> Ten (10) working days with no extension upon receipt of approved NOA and NTP. Failure to comply with the delivery schedule shall mean forfeiture of the contract				
		2. <b>Testing:</b> Prior to acceptance, the COVID Task Force / End user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.				
		3. <b>Training:</b> The supplier shall provide an orientation/ training on the proper use and maintenance of the equipment to the end-users.				
		4. <b>Warranty:</b> Warranty certificate for One (1) year on parts and services. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.				
		5. <b>Manuals:</b> The supplier must provide Manual in English Language (2 hard copies and 1 soft copy of Operator's/User's Manual and 2 hard copies and 1 soft copy of Service/Technical and Maintenance Manual) upon delivery				
		6. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language.				