



Republic of the Philippines
Department of Health
EAST AVENUE MEDICAL CENTER
Hospital Bids and Awards Committee
East Avenue, Quezon City
Philippines



SUPPLEMENTAL BID BULLETIN NO. 1

This **Supplemental Bid Bulletin** shall be applicable for:

1. RA 11469 or Bayanihan to Heal as One Act Emergency Procurement of **Supply, Delivery, Testing and Commissioning of Nasal High Flow Oxygen Therapy Device**

This is being issued to clarify, modify and amend items/specifications in the Technical Specification and Price Quotation.

Emergency Procurement of Supply, Delivery, Testing and Commissioning of Nasal High Flow Oxygen Therapy Device

TECHNICAL SPECIFICATION & PRICE QUOTATION LIST

FROM:

ITEM NO.	ITEM	SPECIFICATIONS	QTY	UNIT	CEILING PRICE PER UNIT
1	Nasal High Flow		1	unit	450,000.00
	<u>Clinical Features:</u>				
	Temperature Settings	31 - 37 C			
	Max Temperature	43 C			
	Humidity Requirement	> 12 - > 33 mg/L (10 - 95% Relative Humidity)			
	Flow Range	10 - 60 Lpm (minimum range)			
	Oxygen Output	60Lpm			
	FiO2 Monitoring Range	21 - 100%			
	FiO2 Monitoring Accuracy	+/- 4%			
	Warm - up Time	at least 10 - 30 mins.			
		(at temp range 31 - 37 C)			
	Alarms	< / = 45 dbA			
	Alarm Silence	At least 1 min.			

“Service to Humanity is the Best Work in Life”



Republic of the Philippines
 Department of Health
EAST AVENUE MEDICAL CENTER
Hospital Bids and Awards Committee
 East Avenue, Quezon City
 Philippines



	<u>Physical</u>				
	Weight	less than 10 lbs.			
	Mobility	with stand and at least 4 wheels			
	<u>Electrical</u>				
	Frequency	50 - 60Hz			
	Current/Voltage Requirement	Autovolt (110 -240)			
	UPS (Uninterrupted Power Supply)	(at least 30 mins to 1 hour)			
	<u>Accessories</u>				
	Breathing Tubes	Disposable			
	Humidification Chamber	Disposable			
	HFNC Interface	Small, Medium, Large			
		Documentary Requirements:			
		1.Certificate of availability of stocks and can deliver within the prescribed delivery schedule.			
		2) Valid and current License to Operate (LTO) as medical device distributor issued by the Philippine Food and Drug Administration. In Case of expired LTO, the following must be submitted: i) Copy of expiredLTP, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO			
	Current/Voltage Requirement	Autovolt (110 -240)			
	UPS (Uninterrupted Power Supply)	(at least 30 mins to 1 hour)			
		Requirements if awarded the Contract:			
		1.Delivery and Completion Schedule: Ten (10) working days with no extension upon receipt of approved NOA and NTP. Failure to comply with the delivery schedule shall mean forfeiture of the contract			



Republic of the Philippines
 Department of Health
EAST AVENUE MEDICAL CENTER
Hospital Bids and Awards Committee
 East Avenue, Quezon City
 Philippines



		2. Testing: Prior to acceptance, the COVID Task Force / End user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.			
		3. Training: The supplier shall provide an orientation/ training on the proper use and maintenance of the equipment to the end-users.			
		4. Warranty: Warranty certificate for One (1) year on parts and services. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.			
		5. Manuals: The supplier must provide Manual in English Language (2 hard copies and 1 soft copy of Operator's/User's Manual and 2 hard copies and 1 soft copy of Service/Technical and Maintenance Manual) upon delivery			
		6. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language.			

TO:

ITEM NO.	ITEM	SPECIFICATIONS	QTY	UNIT	CEILING PRICE PER UNIT
1	Emergency Procurement of Supply, Delivery, Testing, and Commissioning of Nasal High		1	unit	450,000.00

“Service to Humanity is the Best Work in Life”



Republic of the Philippines
 Department of Health
EAST AVENUE MEDICAL CENTER
Hospital Bids and Awards Committee
 East Avenue, Quezon City
 Philippines



	Flow Oxygen Therapy Device				
	<u>Clinical Features</u>				
	Temperature Settings	31 - 37 C (minimum range)			
	Max Temperature	43 C			
	Humidity Requirement	> 12 - > 33 mg/L (10 - 95% Relative Humidity)			
	Flow Range	2 - 60 Lpm (minimum range)			
	Maximum Oxygen Input	60Lpm			
	FiO2 Monitoring Range	21 - 100%			
	FiO2 Monitoring Accuracy	+/- 4%			
	Warm - up Time	at least 10 - 30 mins. (at temp range 31 - 37 C)			
	Auditory Alarms	>/ = 45 dbA;			
	Visual Alarms	can detect: unplugged breathing tube; leak in the system; blockage in the system; if measured oxygen level has fallen below allowed limit; if measured oxygen has exceeded the allowed limit; if unit cannot reach the target flow setting; if unit cannot reach target temperature setting; if unit has been disconnected from the main power supply			
	Alarm Silence	At least 1 min.			
	Disinfection kit	with Disinfection Filter, Sponge			
	Oxygen analyzer accuracy	<+/- (2.5% + 2.5% of gas level)			
	<u>Physical</u>				
	Weight	less than 10 lbs.			
	Mobility	with hospital stand and at least 4 wheels, and pole mounting tray			
	<u>Electrical</u>				
	Frequency	50 - 60Hz			
	Current/Voltage Requirement	Autovolt (110 -240)			



Republic of the Philippines
 Department of Health
EAST AVENUE MEDICAL CENTER
Hospital Bids and Awards Committee
 East Avenue, Quezon City
 Philippines



	UPS (Uninterrupted Power Supply)	(at least 30 mins to 1 hour)			
	Accessories				
	Heated Non-condensing Breathing Tubes/Circuit	Disposable			
	Humidification Chamber	Disposable			
	HF Nasal Cannula Patient Interface	Small, Medium, Large, Tracheostomy connection adapter, Mask interface adapter			
	Air Filter	3 months maximum period of use			
	Miscellaneous				
	Oxygen inlet extension kit				
	Hospital stand	with wheels			
	Pole mounting tray				
		Documentary Requirements:			
		1.Certificate of availability of stocks and can deliver within the prescribed delivery schedule.			
		2) Valid and current License to Operate (LTO) as medical device distributor issued by the Philippine Food and Drug Administration. In Case of expired LTO, the following must be submitted: i) Copy of expiredLTP, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO			
		Requirements if awarded the Contract:			
		1.Delivery and Completion Schedule: Ten (10) working days with no extension upon receipt of approved NOA and NTP. Failure to comply with the delivery schedule shall mean forfeiture of the contract			
		2. Testing: Prior to acceptance, the COVID Task Force / End user			



Republic of the Philippines
 Department of Health
EAST AVENUE MEDICAL CENTER
Hospital Bids and Awards Committee
 East Avenue, Quezon City
 Philippines



		shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.			
		3. Training: The supplier shall provide an orientation/ training on the proper use and maintenance of the equipment to the end-users.			
		4. Warranty: Warranty certificate for One (1) year on parts and services. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.			
		5. Manuals: The supplier must provide Manual in English Language (2 hard copies and 1 soft copy of Operator's/User's Manual and 2 hard copies and 1 soft copy of Service/Technical and Maintenance Manual) upon delivery			
		6. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language.			

This Supplemental Bid Bulletin shall form part of the Technical Specification and Price Quotation. Any provisions in the Technical Specification and Price Quotation inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 3rd day of June, 2020 in Quezon City.

JOSE C. CALIXTRO, MPA
Chief Administrative Officer
(SGD)

“Service to Humanity is the Best Work in Life”