

Section VII. Technical Specifications

Technical Specifications

Item	Specification	Statement of Compliance
		<i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>

ITEM NO.	ITEMS	SPECIFICATIONS	UNIT	STATEMENT OF COMPLIANCE
1	PFT Filter mouthpiece	round, recommended to be replaced per patient (compatible with Pulmonary Function Test Machine, Brand: Cosmed Q-Box)	piece	
2	PFT Filter mouthpiece	elliptical, recommended to be replaced per patient (compatible with Pulmonary Function Test Machine, Brand: Cosmed Q-Box)	piece	
3	Biopsy forcep	compatible with the existing bronchoscope, OD 1.8 * 1800mm with spike, coated 2.8 working channel innovation 4 bar linkage design for excellent rigidity, smooth jaws, tiny interval between jaws 15 units/box	box	

4	Biopsy forcep	compatible with the existing bronchoscope, OD 1.8 * 1800mm without spike, coated 2.8 working channel innovation 4 bar linkage design for excellent rigidity, smooth jaws, tiny interval between jaws 15 units/box	box	
5	Cytology brush	compatible with the existing bronchoscope, disposable, OD, 2.3 * 2300mm, U shaped 2.9 working channel innovation brush design with no risk of drop off, both straight shaped brush, U shaped brush 15 units/ box	box	
6	Spray catheter	compatible with the existing bronchoscope, disposable, OD, 2.3* 2300mm, direct injection , A type and B type, wrench proof design, Large spraying range and uniform spray deposition 15 units/ box	box	
7	Nose clips	100 pieces per bag	bag	
8	CO2 absorber cartridge PN: CO3540-01-08	disposable, non-refillable	piece	
9	Calibration gas, Compressed gas, N.O.S.	5% CO2, 16% O2, balance N2, without acetylene, H Tank	cylinder	
10	Gas, DLCO mix	CO 0.3%, CH4 0.3%, O2 21%, bal. N2, without acetylene H Tank	cylinder	
11	Adhesive bandage with compression stanching pad	<ul style="list-style-type: none"> - wound site protection - absorbable hemostatic pressure dressing - sterile - pad, 6mm thick pad excellent in blood absorbancy - 27mm longitudinal width oblong covers the gap of puncture site on the skin and that of blood vessel and gives the astriction on the puncture site - oblong shape - urethane non woven fabric type excellent in elasticity & flexibility - prevent the gap of pad by wrapping up the pad from 4 direction slightly extending the tape - tape are: 39mm x 80mm - pad are: 27mm x 15mm - 50 sheets per box 	box	
12	Bag-valve mask resuscitator	bag-valve mask resuscitator, adult, 1.5L, size 5 mask, with peep valve	piece	
13	Bronchoscope adaptor	reduces gas leak during bronchoscopy and provides swivel help to reduce tension on Endotracheal Tube connections, single patient use ,	piece	
14	C02210-02-08 Permapure line	good for up to 200 PFT use (compatible with Pulmonary Function Test Machine, Brand: Cosmed Q-Box)	cylinder	
15	Flex tube connector	isothermal adult flex connector, placed between the endotracheal tube of the patient and a heat/ moisture filter or a humidifying filter they allow to avoid excessive tension on the endotracheal tube	piece	

16	Mask	tracheostomy mask for adult use with swivel adaptor and accept 22-mm corrugated tubing, also the swivel connector permits access from either side of the patient, with strap clips allow the mask to be removed with minimal disturbance to the patient	piece	
----	------	---	-------	--

TERMS & CONDITIONS:

1. The quantities specified are estimated quantities only and may be decreased based on the actual need of the hospital and availability of fund. It is understood that EAMC is not bound to purchase all of the quantities/items called for in the Invitation to Bid (IB).
2. Deliveries shall have at least eighteen (18) months expiration date/shelf-life. For items with shorter expiration date, the Supplier shall submit an assurance letter prior to deliver for approval of the Head of the Procuring Entity (HOPE) or his authorized representative.

<i>Conforme:</i>
<i>Company/Bidder's Name:</i>
<i>Name and Signature of Authorized Representative:</i>
<i>Date:</i>