

Section VII. Technical Specifications

Technical Specifications

Item	Specification	Statement of Compliance
		<i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i>

ITEM NO.	ITEMS	SPECIFICATIONS	UNIT	STATEMENT OF COMPLIANCE
1	Biological Indicator	for steam sterilization, 4ml, 50 vials/box (winning bidder shall provide 5 machine)	box	
2	Bowie Dick test	pre-packed, non-woven wrap / test card	piece	
3	Chemical indicator strip	with objective reading result, does not require color interpretation, provides instant result at a glance	box	
4	CTT (Chest Tube Thoracostomy) bottle	2 way, glass, with stopper, 1 liter capacity	bottle	
5	Industrial salt	for autoclave, 50 kilos per sack	sack	

ITEM NO.	ITEMS	SPECIFICATIONS	UNIT	STATEMENT OF COMPLIANCE
6	Labelling sticker	29x28mm labels with double adhesive support, with process indicator for steam sterilization process, must be Class 1 (ISO 11140-1:2009) Label material: vellum paper, Label adhesive material: acrylic non-toxic and must have color-change indicator, label must at least conform to EN ISO 9001:2008. Winning bidder shall provide one (3) medical-grade labelling device. Heavy duty and durable, light weight. With at least 3-4 lines that can print lot number, sterilizer code, batch number, operator and the expiry date for full documentation of information traceability. With alpha-numeric and other symbols of at least 12-15, incapable of alteration of information or smudging of print after exposure to sterilization process	roll	
7	Pouch	sterile packaging reel, gussetted, 30cm x 100m, at least 60-70 grams per square meter (gsm), with steam, formaldehyde and ethylene oxide indicators, winning bidder shall provide 2 foot sealer / automated sealer	roll	
8	Pouch	sterile packaging reel, gussetted, 20cm x 100m, at least 60-70 grams per square meter (gsm), with steam, formaldehyde and ethylene oxide indicators, winning bidder shall provide 2 foot sealer / automated sealer	roll	
9	Pouch	sterile packaging reel gussetted, 15cm x 100m, at least 60-70 grams per square meter (gsm), with steam, formaldehyde and ethylene oxide indicators, winning bidder shall provide 1 foot sealer / automated sealer	roll	
10	Pouch	sterile packaging reel gussetted, 10cm x 100m, at least 60-70 grams per square meter (gsm), with steam, formaldehyde and ethylene oxide indicators, winning bidder shall provide 1 foot sealer / automated sealer	roll	
11	Tape	autoclave, lead free, solvent free, 3/4 size	roll	
12	Water sediment filter	thread	piece	
13	Water sediment filter	ceramic	piece	

TERMS & CONDITIONS:

1. The quantities specified are the estimated quantities only and may be decreased based on the actual need of the hospital and availability of fund. It is understood that EAMC is not bound to purchase all of the quantities/items called for in Invitation to Bid (ITB).

2. Deliveries shall have at least eighteen (18) months expiration date/shelf-life. For items with shorter expiration date, the Supplier shall submit an assurance letter prior to deliver for approval of the Head of the Procuring Entity (HOPE) or his authorized representative.

<i>Conforme:</i>
<i>Company/Bidder's Name:</i>
<i>Name and Signature of Authorized Representative:</i>
<i>Date:</i>